

JAN-03-2001 11:31 AM CHARLES HOFFMAN


520 443 0570

P.04

Sent by: CAHILL TON

6020564208;

01/03/01 11:05AM; Jan 3, 2001 #844; Page 4/5

**IN THE UNITED STATES PATENT AND TRADEMARK OFFICE**

EXAMINER: B. Yarnell Docket No.: 5899-A-01  
APPLICANT: G Gobeli Art Unit: 3739  
SERIAL NO.: 08/873,999 Date: January 3, 2001  
FILED: June 12, 1997  
FOR: METHOD AND DEVICE FOR GLUCOSE CONCENTRATION  
MEASUREMENT WITH SPECIAL ATTENTION TO BLOOD  
GLUCOSE DETERMINATIONS

**AMENDMENT**

Commissioner for Patents

Washington, D.C. 20231

Dear Sir:

In response to the office action mailed 07/07/2000, please amend the above application as follows:

Please amend the claims as follows:

In Claim 38, line 7, before "quarter", delete "the", substitute --a--.

Please amend the specification as follows:

I hereby certify that this correspondence is being deposited with the United States Postal Service as first class mail in an envelope addressed to: Commissioner for Patents, Washington, D.C. 20231	
January 3, 2001	
Name of Applicant, Inventor or Registered Rep. <u>Charles K. Hoffman</u>	
Signature <u>Charles K. Hoffman</u>	Date <u>1/3/01</u>

Page 3, line 20, delete "sucrose", substitute --glucose--.

Page 5, line 14, delete "53", substitute --57--; delete "7", substitute --3--.

Page 8, line 18, after "based", insert --on--.

### REMARKS

A request for a three-month extension of time is enclosed, accompanied by the statutory fee of \$445.00.

In the Office Action, the examiner rejected Claims 1-9, 11-24, 27-31, 33-35 and 38-42 under 35 U.S.C. 103 (a) as a patentable over Cote et al. in view of Pezzaniti et al. and Kuhls.

The rejection is respectfully traversed, in view of the following arguments.

Claim 1 recites:

"introducing the beam into the anterior chamber such that it is refracted within the anterior chamber, impinges on the iris, is reflected therefrom, and then exits the anterior chamber approximately collinear with the beam immediately before the point at which it is introduced into the anterior chamber".

It is respectfully submitted that the foregoing limitation of Claim 1 is clearly illustrated in Fig. 8A, wherein the external "probe beam" 55A is approximately parallel to the iris 56 as it (the beam 55A) is introduced into the cornea 57, and then is refracted as indicated by 55B to a spot 58